



Llywodraeth Cymru  
Welsh Government

# Medical Device Alert

**07 August 2018**

Parc Cathays  
Caerdydd CF10 3NQ

Cathays Park  
Cardiff CF10 3NQ

**Orthopaedic bone plates and cortical screws: ADVANSYS MLP-DLP; ADVANSYS TTC; Large QWIX; TIBIAXYS and UNI-CP–Sterile – Risk of infection**

1. Identify and quarantine all affected devices.
2. Follow actions recommended in the manufacturer's Field Safety Notice.
3. Complete the certificate of acknowledgment attached to the Field Safety Notice and return it to the manufacturer.
4. Report all adverse events involving this device to Newdeal SAS and the MHRA or the appropriate Devolved Administration.

**Sent to:**

NHS Wales Shared Services Partnership  
Directors of Public Health,  
Chief Executives, NHS Trusts  
Chief Executives, LHBs  
Medical Directors, NHS Trusts  
Nurse Directors, LHBs  
Executive Nurses, NHS Trusts  
NHS Direct  
CSSIW  
Social Services  
HIW

A list of people who need to have early sight of this information is given in the Notice.  
This Notice has been endorsed by the Welsh Government as being relevant to NHS Wales.

## **Points of particular importance in Wales:**

**Healthcare Quality Division  
Welsh Government  
Cathays Park  
Cardiff CF10 3NQ**

This document is available online at:  
<https://www2.nphs.wales.nhs.uk/contacts.nsf>

Ffôn/Tel: 029 2082 3624/029 2082 5510  
Ebost/Email: [Haz-Aic@wales.gsi.gov.uk](mailto:Haz-Aic@wales.gsi.gov.uk)



# Medical Device Alert

MDA/2018/028

Issued: 01 August 2018 at 14:00

Valid until: August 2019

Orthopaedic bone plates and cortical screws: ADVANSYS MLP-DLP; ADVANSYS TTC; Large QWIX; TIBIAXYS and UNI-CP–Sterile – Risk of infection

## Summary

Manufactured by Newdeal SAS – Risk of infection from compromised packaging.

## Action

1. Identify and quarantine all affected devices.
2. Follow actions recommended in the manufacturer's [Field Safety Notice](#).
3. Complete the certificate of acknowledgment attached to the Field Safety Notice and return it to the manufacturer.
4. Report all adverse events involving this device to Newdeal SAS and the MHRA or the appropriate Devolved Administration.

### Action by

All users of the affected medical device.

### Deadlines for actions

Actions underway: 15 August 2018

Actions complete: 29 August 2018

**Medical Device Safety Officers** (in England): ask the manufacturer to add you to their distribution list for field safety notices (FSNs). This is to help with reconciliation.

**Remember:** if your organisation receives an [FSN](#) from a manufacturer, always act on it. **Do not wait** for a communication from MHRA.

## Device details

The affected devices are used in various bone and joint reconstruction procedures.

In addition to the Field Safety Notice which details affected product, please refer to the spreadsheet which accompanies this MDA for additional unique device identification information.

## Problem / background

In May 2018, Newdeal SAS, a company within Integra LifeSciences Group, issued a Field Safety Notice informing distributors and clinicians of devices possibly affected with defective packaging. This defect may result in insufficient sealing, a potential consequence of which is an increased risk of infection.

This Medical Device Alert is being issued to ensure that all hospitals are aware of the issue and that adequate action is taken to mitigate potential risk to patients.

## Manufacturer contacts

Newdeal SAS

Tel: +33 (0)4 37 47 51 51

Email: [emea-fsca-recon@integralife.com](mailto:emea-fsca-recon@integralife.com) or [marilyse.latour@integralife.com](mailto:marilyse.latour@integralife.com)

Fax/telecopy: +33 (0)4 37 47 59 30

## Distribution

If you are responsible for cascading these alerts in your organisation, these are our suggested distribution lists.

### **Trusts (NHS boards in Scotland)**

CAS and NICAS liaison officers for onward distribution to all relevant staff including:

- Orthopaedic surgeons
- Sterile services departments
- Supplies managers
- Theatre managers
- Theatres

### ***Independent distribution***

#### **Establishments registered with the Care Quality Commission (CQC) (England only)**

- Hospitals in the independent sector
- Independent treatment centres

Please note: CQC and OFSTED do not distribute these alerts. Independent healthcare providers and social care providers can sign up to receive MDAs directly from the Central Alerting System (CAS) by sending an email to: [safetyalerts@mhra.gov.uk](mailto:safetyalerts@mhra.gov.uk) and requesting this facility.

## Enquiries

### England

Send enquiries about this notice to MHRA, quoting reference number **MDA/2018/028** or **2018/005/025/291/009**.

### Technical aspects

Hasan Samee-Ahmed, MHRA

Tel: 020 3080 6807

Email: [hasan.samee-ahmed@mhra.gov.uk](mailto:hasan.samee-ahmed@mhra.gov.uk)

### Clinical aspects

Devices Clinical Team, MHRA

Tel: 020 3080 7274

Email: [dct@mhra.gov.uk](mailto:dct@mhra.gov.uk)

### Reporting adverse incidents in England

Through Yellow Card <https://yellowcard.mhra.gov.uk/>

### Northern Ireland

Alerts in Northern Ireland are distributed via the [NICAS system](#).

Enquiries and adverse incident reports in Northern Ireland should be addressed to:

Northern Ireland Adverse Incident Centre, CMO Group,  
Department of Health, Social Services and Public Safety

Tel: 028 9052 3868

Email: [niaic@health-ni.gov.uk](mailto:niaic@health-ni.gov.uk)  
<https://www.health-ni.gov.uk/niaic>

### Reporting adverse incidents in Northern Ireland

Please report directly to NIAIC using the [forms on our website](#).

### Scotland

Enquiries and adverse incident reports in Scotland should be addressed to:

Incident Reporting and Investigation Centre, Health Facilities Scotland, NHS National Services Scotland

Tel: 0131 275 7575

Email: [nss.irc@nhs.net](mailto:nss.irc@nhs.net)

### Reporting adverse incidents in Scotland

NHS Boards and Local Authorities in Scotland – report to [Health Facilities Scotland](#).

Contractors such as private hospitals carrying out NHS work and private care homes that accept social work funded clients – report to [Health Facilities Scotland](#).

Private facilities providing care to private clients report to the [Care Inspectorate](#) and [MHRA](#).

## **Wales**

Enquiries in Wales should be addressed to:  
Healthcare Quality Division, Welsh Government

Tel: 02920 823 624 / 02920 825 510

Email: [Haz-Aic@wales.gsi.gov.uk](mailto:Haz-Aic@wales.gsi.gov.uk)

### **Reporting adverse incidents in Wales**

Report to MHRA through Yellow Card <https://yellowcard.mhra.gov.uk/> and follow specific advice for reporting in Wales in [MDA/2004/054 \(Wales\)](#).

MHRA is a centre of the Medicines and Healthcare products Regulatory Agency, an executive agency of the Department of Health and Social Care.

© Crown Copyright 2018

Addressees may take copies for distribution within their own organisations